

Ref No: S123176

Certificate of Conformance

Issued to: Valley Safety Services Assoc, Inc.

1029 Pine St.

Palmer, MA 01069

Equipment Description	Model	S/N	
Accu-Gold Digitizer Module	AGDM	40-0345	
Accu-Gold Multi-Sensor	AGMS-D	41-0436	
Accu-Gold Light Sensor	AGLS	01-0137	
Accu-Gold Multi-Sensor	AGMS-M	42-0180	

The equipment identified above has been calibrated and tested using Radcal calibration and acceptance procedure PP102, Radcal Quality Manual PP1007, Radcal Policy and Procedure PP1038, Pl1045, Pl1055 and other related documents. The equipment has been found to conform in all respects. These test procedures are designed to ensure that the tested equipment meets or exceeds all aspects of Radcal's published product specifications and requirements. Radcal is an ACLASS accredited calibration lab that meets the requirements of ISO 17025 and ANSI/NCLS Z540-1, cert number AC-1553.

All measurements performed during the testing employ equipment traceable to NIST or another recognized National Laboratory such as Physikalisch-Technische Bundesansalt (PTB).

For additional information please refer to Radcal's Product note: "The Importance of Conformance Testing". Radcal recommends revalidation in 12 months.

Certificate Issue Date 20-Feb-18

Authorized Representative

Service Report

Ref No: S123176

Received: 12-Feb-18

The equipment was tested for conformance with Radcal specifications using applicable Conformance test procedures. These procedures include inspection, operation with an x-ray machine and electrical test. The results are summarized below:

Model Number	Serial Number	Description	Meets Mfr Spec	Spec limit (±)	Cal Date
AGDM	40-0345	Accu-Gold Digitizer Module	Yes	Pass/Fail	20-Feb-18
AGLS	01-0137	Accu-Gold Light Sensor	Yes	Pass/Fail	20-Feb-18
AGMS-D	41-0436	Accu-Gold Multi-Sensor	Yes	Pass/Fail	20-Feb-18
AGMS-M	42-0180	Accu-Gold Multi-Sensor	Yes	Pass/Fail	20-Feb-18

Service requested:

Perform conformance test, inspection and issue certificate.

Service performed:

Upon receipt, the equipment met manufacturer's specifications. AGMS-D was calibrated for optimal performance. AGMS-M have WAI calibration added. Issued Certificate of Conformance.