



Ref No: S124817

Certificate of Conformance

Issued to: Upstate Medical Physics
1290 Blossom Dr.
Victor, NY 14564


<u>Equipment Description</u>	<u>Model</u>	<u>S/N</u>
Accu-Gold Digitizer Module	AGDM+	48-0692
Accu-Gold Light Sensor	AGLS	01-0219
Accu-Gold Multi-Sensor	AGMS-D+	43-0598

The equipment identified above has been calibrated and tested using Radcal calibration and acceptance procedure PP1102, Radcal Quality Manual PP1007, Radcal Policy and Procedure PP1038, PI1045, PI1055 and other related documents. The equipment has been found to conform in all respects. These test procedures are designed to ensure that the tested equipment meets or exceeds all aspects of Radcal's published product specifications and requirements. Radcal is an ACLASS accredited calibration lab that meets the requirements of ISO 17025 and ANSI/NCLZ Z540-1, cert number AC-1553.

All measurements performed during the testing employ equipment traceable to NIST or another recognized National Laboratory such as Physikalisch-Technische Bundesanstalt (PTB).

For additional information please refer to Radcal's Product note: "The Importance of Conformance Testing". Radcal recommends revalidation in 12 months.

Certificate Issue Date 28-Nov-18

By: 
Authorized Representative



Ref No: S123351

Certificate of Conformance

Issued to: Upstate Medical Physics
1290 Blossom Dr.
Victor, NY 14564


<u>Equipment Description</u>	<u>Model</u>	<u>S/N</u>
Accu-Gold Digitizer Module	AGDM+	48-0692
Accu-Gold Multi-Sensor	AGMS-D+	43-0598

The equipment identified above has been calibrated and tested using Radcal calibration and acceptance procedure PP1102, Radcal Quality Manual PP1007, Radcal Policy and Procedure PP1038, PI1045, PI1055 and other related documents. The equipment has been found to conform in all respects. These test procedures are designed to ensure that the tested equipment meets or exceeds all aspects of Radcal's published product specifications and requirements. Radcal is an ACLASS accredited calibration lab that meets the requirements of ISO 17025 and ANSI/NCLS Z540-1, cert number AC-1553.

All measurements performed during the testing employ equipment traceable to NIST or another recognized National Laboratory such as Physikalisch-Technische Bundesanstalt (PTB).

For additional information please refer to Radcal's Product note: "The Importance of Conformance Testing". Radcal recommends revalidation in 12 months.

Certificate Issue Date 04-Apr-18

By: 
Authorized Representative

Radcal Corporation
426 W. Duarte Rd. Monrovia, CA 91016
Tel: (626) 357-7921 FAX: (626) 357-8863 email: service@radcal.com

This Certificate of Conformance shall not be reproduced except in full, without the written approval of Radcal Corporation.