

Certificate number: o60209451i148204-A-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product Xi R/F & MAM
Serial number 148204
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 1/7/2022
As left 2/1/2022
Adjustment done Yes
Tested by Amber Ramos
Channell Kennedy
Henzelo Barner
Keosha Lloyd
Lloynesha Baker

Approved by


Josefina Diaz
Lab Technician

Certificate date 2/4/2022

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in compliance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Time: 0.5 %

Voltage (R/F): 2 %

Voltage (MAM): 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60209451i148204-B-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product Xi R/F & MAM
Serial number 148204
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 1/7/2022
As left 2/1/2022
Adjustment done Yes
Tested by Amber Ramos
Channell Kennedy
Henzelo Barner
Keosha Lloyd
Lloynesha Baker

Approved by


Josefina Diaz
Lab Technician

Certificate date

2/4/2022



Certificate number: o60209806i205689-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product Xi Survey
Serial number 205689
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 1/18/2022
As left 1/21/2022
Adjustment done Yes
Tested by Amber Ramos
Channell Kennedy

Approved by


Josefina Diaz
Lab Technician

Certificate date 2/3/2022

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in compliance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications
Air kerma: 10 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60209940i150112-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product Xi CT
Serial number 150112
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 1/20/2022
As left 1/21/2022
Adjustment done Yes
Tested by Amber Ramos

Approved by


Josefina Diaz
Lab Technician

Certificate date 2/3/2022

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in compliance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Energy dependence: < 5 %
Radial uniformity: 2 %, Axial uniformity: 3 %
Humidity: < 0.3 %, Temp. and pressure: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60209669i146608-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product Xi Light
Serial number 146608
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 1/13/2022
As left 1/18/2022
Adjustment done Yes
Tested by Channell Kennedy
Joseph Hertl

Approved by


Josefina Diaz
Lab Technician

Certificate date 2/3/2022

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in compliance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications

Illuminance: 3 %

Luminance: 3 %

