


Certificate number: o60213068i263036-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Survey
Serial number 263036
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 4/6/2022
As left 4/6/2022
Adjustment done Yes
Tested by Joseph Hertl

Approved by

Claudia Castillo

Claudia Castillo
Finalization operator
4/8/2022

Certificate date

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in compliance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer

LANDAUER MEDICAL PHYSICS
ELLINGTON CT 205
13031 CAMILLUS
US

Laboratory

Unfors RaySafe AB
Uggledalsvägen 29
SE-427 40 Billdal
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com


CUSTOMER INSTRUMENT

Product X2 CT
Serial number 293251
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2022-03-31
Adjustment done Yes
Tested by André Gillberg

Approved by



Philip Kinnander
Lab Technician

Certificate date

2022-04-12

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60213016i272670-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Light
Serial number 272670
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 4/4/2022
As left 4/4/2022
Adjustment done Yes
Tested by Channell Kennedy

Approved by

Claudia Castillo
Finalization operator
4/11/2022

Certificate date

CALIBRATION RESULT

AS FOUND: One or more measured and tested values of this certificate are outside of the specifications by a margin less than the expanded measurement uncertainty. Therefore, it is not possible to state non-compliance with specifications for those measurements.


AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications

Illuminance: 3 %
Luminance: 3 %



Certificate number: o60213070i279216-A-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 279216
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 4/6/2022
As left 4/7/2022
Adjustment done Yes
Tested by Henzelo Barner
Keosha Lloyd
Kierra Sullivan

Approved by

Claudia Castillo
Finalization operator

Certificate date 4/8/2022

CALIBRATION RESULT

AS FOUND: One or more measured and tested values of this certificate are within the specifications by a margin less than the expanded measurement uncertainty. Therefore, it is not possible to state compliance with specifications for those measurements.

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.


Relevant instrument specifications

Air Kerma: 5 %
Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60213070i279216-B-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 279216
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 4/6/2022
As left 4/7/2022
Adjustment done Yes
Tested by Henzelo Barner
Keosha Lloyd
Kierra Sullivan

Approved by


A handwritten signature in blue ink that reads "Claudia Castillo".

Claudia Castillo
Finalization operator
4/8/2022

Certificate date



Certificate number: o60213042i279575-B-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 279575
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 4/5/2022
As left 4/5/2022
Adjustment done Yes
Tested by Joseph Hertl

Approved by

Claudia Castillo

Claudia Castillo
Finalization operator
4/8/2022

Certificate date



Certificate number: o60213042i279575-A-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com


CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 279575
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 4/5/2022
As left 4/5/2022
Adjustment done Yes
Tested by Joseph Hertl

Approved by


Claudia Castillo
Finalization operator
4/8/2022

Certificate date

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in compliance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in compliance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

