

# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe AB  
Uggledalsvägen 29  
SE-427 40 Billdal  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com


## CUSTOMER INSTRUMENT

Product X2 Survey  
Serial number 282096  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found Not performed  
As left 2023-02-13  
Adjustment done Yes  
Tested by Adam Eckebrant

Approved by



Örjan Arnström  
Finalization operator  
2023-02-16

Certificate date

## CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Certificate of Calibration

UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Product:	RaySafe X2
Model:	Light Sensor
Serial Number:	300067

**Calibration Certificate:**  
**Calibration Equipment List**

Date: 2022-12-19

Your instrument has passed several strict tests and has been approved for delivery.  
Enclosed you will find detailed documents for your instrument.

This certificate may not be reproduced other than in full, except with the prior written approval of  
Unfors RaySafe



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe AB  
Uggledalsvägen 29  
SE-427 40 Billdal  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 R/F  
Serial number 311361  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found Not performed  
As left 2023-01-23  
Adjustment done Yes  
Tested by Kreshnik Raca

Approved by



Örjan Arnström  
Finalization operator

Certificate date

2023-02-16

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe AB  
Uggledalsvägen 29  
SE-427 40 Billdal  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com


## CUSTOMER INSTRUMENT

Product X2 R/F  
Serial number 311361  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found Not performed  
As left 2023-01-23  
Adjustment done Yes  
Tested by Kreshnik Raca

Approved by



Örjan Arnström  
Finalization operator  
2023-02-16

Certificate date

## CALIBRATION RESULT

**AS LEFT:** All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %  
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.





# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe AB  
Uggledalsvägen 29  
SE-427 40 Billdal  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 CT  
Serial number 308613  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found Not performed  
As left 2023-01-27  
Adjustment done Yes  
Tested by André Gillberg

Approved by



Örjan Arnström  
Finalization operator  
2023-02-16

Certificate date

## CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe AB  
Uggledalsvägen 29  
SE-427 40 Billdal  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com


## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 311863  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found Not performed  
As left 2023-01-13  
Adjustment done Yes  
Tested by Kreshnik Raca

Approved by



Örjan Arnström  
Finalization operator  
2023-02-16

Certificate date

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe AB  
Uggledalsvägen 29  
SE-427 40 Billdal  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com


## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 311863  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found Not performed  
As left 2023-01-13  
Adjustment done Yes  
Tested by Kreshnik Raca

Approved by



Örjan Arnström  
Finalization operator  
2023-02-16

Certificate date

## CALIBRATION RESULT

**AS LEFT:** All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air Kerma: 5 %

Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

