

Customer: Jen Newtown
Upstate Medical Physics
1290 Blossom Drive
Victor, NY 14564

Instrument
Victoreen Model 451B

Serial Number
452

Precision Check				
Test 1	Test 2	Test 3	Mean	Results
106.00 mR/hr	107.00 mR/hr	107.00 mR/hr	106.67 mR/hr	Satisfactory

Accuracy Check			
Range	Target Value	As Found	As Left
Rate	40 R/hr	37 R/hr	37 R/hr
Rate	10 R/hr	10.5 R/hr	10.5 R/hr
Rate	4 R/hr	4.1 R/hr	4.1 R/hr
Rate	1 R/hr	1.08 R/hr	1.08 R/hr
Rate	400 mR/hr	400 mR/hr	400 mR/hr
Rate	100 mR/hr	107 mR/hr	107 mR/hr
Rate	40 mR/hr	38 mR/hr	38 mR/hr
Rate	10 mR/hr	9.8 mR/hr	9.8 mR/hr
Rate	4 mR/hr	3.8 mR/hr	3.8 mR/hr
Rate	1 mR/hr	1.07 mR/hr	1.07 mR/hr
Integrate	200 mR	200 mR	200 mR
Integrate	50 mR	53 mR	53 mR

Readings with * indicate ranges where As-Found readings are >20% of Target value. Readings with ** indicate As-left readings are >10.00% of Target value

Outer Physical Check: <i>Pass</i>	
Internal Check: <i>Pass</i>	Tap Test: <i>Pass</i>
	Desiccant Check: <i>Pass</i>

Comments: All readings higher than 40 mR/hr were obtained using the J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N: 9141). All other readings were obtained using the Tech Ops Model 773 Cs-137 Beam Calibrator (S/N: S-1110). As-Found/As-Left Calibration Factors: F1 = 96/96, F2 = 92/92, F3 = 105/105, F4 = 102/102, F5 = 95/95, INT = 100/100 Dried/replaced desiccant.

Chris Hartman
Calibration
Technician

QA
Review:

Calibration Date: 09/07/2023
Expires: 09/07/2024

Atmospheric Conditions - Temperature: 73°F Humidity: 48% Barometric Pressure: 29.71"hg

This calibration was performed by RSCS using one or more of the following NIST Traceable radiation sources:

Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0%.

J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley

Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7%.

J.L. Shepherd and Associates Model 81-12 Cs-137 Dual Source Gamma Calibrator (SN 7145), characterized using Exradin Models A6 (S/N 185) / A4 (S/N 220) / A5 (S/N 261), GE RRS-131 (S/N 95100178), and Standard Imaging Electrometer Model Maxx4000 (S/N J171510) in accordance with methods specified in RSCS TSD 20-023, with estimated uncertainty of 6.7%.

RSCS Model 149 Neutron Calibrator, Pu:Be Source Model MRC Model 2726-B (S/N 495), characterized using Far West Technologies Model Rem500 TEPC (S/N 271) in accordance with the methods specified in RSCS TSD 21-049, with an uncertainty of 9.4% at a 95% confidence interval.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCCL Z540-1-1994

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2005 while demonstrating technical competence in the field of calibration. Refer to the Scope of Accreditation AC-2079 for information on the types of calibrations to which this accreditation applies.

This calibration certificate shall not be reproduced except in full without the express written consent of RSCS, Inc.

Customer: Jen Newtown
Upstate Medical Physics
1290 Blossom Drive
Victor, NY 14564

Instrument
Victoreen Model 451B

Serial Number
1074

Precision Check				
Test 1	Test 2	Test 3	Mean	Results
10.10 mR/hr	10.20 mR/hr	10.10 mR/hr	10.13 mR/hr	Satisfactory

Accuracy Check			
Range	Target Value	As Found	As Left
Rate	40 R/hr	38 R/hr	37 R/hr
Rate	10 R/hr	11.2 R/hr	10.8 R/hr
Rate	4 R/hr	4.1 R/hr	4.1 R/hr
Rate	1 R/hr	1.08 R/hr	1.08 R/hr
Rate	400 mR/hr	430 mR/hr	390 mR/hr
Rate	100 mR/hr	112 mR/hr	101 mR/hr
Rate	40 mR/hr	39 mR/hr	39 mR/hr
Rate	10 mR/hr	10.1 mR/hr	10.1 mR/hr
Rate	4 mR/hr	4.2 mR/hr	4.2 mR/hr
Rate	1 mR/hr	0.92 mR/hr	0.92 mR/hr
Integrate	200 mR	200 mR	200 mR
Integrate	50 mR	53 mR	53 mR

Readings with * indicate ranges where As-Found readings are >20% of Target value. Readings with ** indicate As-left readings are >10.00% of Target value

Outer Physical Check: <i>Pass</i>	
Internal Check: <i>Pass</i>	Tap Test: <i>Pass</i>
	Desiccant Check: <i>Pass</i>

Comments: All readings 100 mR/hr or higher were obtained using the J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N: 9141). All other readings were obtained using the Tech Ops Model 773 Cs-137 Beam Calibrator (S/N: S-1110). As-Found/As-Left Calibration Factors = F1: 96/96, F2: 93/93, F3: 105/97, F4: 99/99, F5: 96/93, Int 100/(100/100).

Mark Nelson
Calibration
Technician



QA
Review:



Calibration Date: 07/18/2023
Expires: 07/18/2024

JUL 19 2023

Atmospheric Conditions - Temperature: 73°F Humidity: 46% Barometric Pressure: 29.89"hg

This calibration was performed by RSCS using one or more of the following NIST Traceable radiation sources:

Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0%.

J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7%.

J.L. Shepherd and Associates Model 81-12 Cs-137 Dual Source Gamma Calibrator (S/N 7145), characterized using Exradin Models A6 (S/N 185) / A4 (S/N 220) / A5 (S/N 261), GE RSS-131 (S/N 95100178), and Standard Imaging Electrometer Model Max4000 (S/N J171510) in accordance with methods specified in RSCS TSD 20-023, with estimated uncertainty of 6.7%.

RSCS Model 149 Neutron Calibrator, Pu:Be Source Model MRC Model 2726-B (S/N 495), characterized using Far West Technologies Model Rem500 TEPC (S/N 271) in accordance with the methods specified in RSCS TSD 21-049, with an uncertainty of 9.4% at a 95% confidence interval.

The reported expanded uncertainty of measurement is stated as the standard uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement.

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2005 while demonstrating technical competence in the field of calibration. Refer to the Scope of Accreditation AC-2079 for information on the types of calibrations to which this accreditation applies.

This calibration certificate shall not be reproduced except in full without the express written consent of RSCS, Inc.

Customer:

Upstate Medical Physics
1290 Blossom Drive
Victor, NY 14564

Instrument

Fluke Biomedical Model 451P

Serial Number

3096

Precision Check				
Test 1	Test 2	Test 3	Mean	Results
10.10 mR/hr	10.30 mR/hr	9.80 mR/hr	10.07 mR/hr	Satisfactory

Accuracy Check			
Range	Target Value	As Found	As Left
Rate	4 R/hr	3.4 R/hr	4.1 R/hr
Rate	1 R/hr	0.84 R/hr	0.99 R/hr
Rate	400 mR/hr	340 mR/hr	400 mR/hr
Rate	100 mR/hr	88 mR/hr	104 mR/hr
Rate	40 mR/hr	37 mR/hr	39 mR/hr
Rate	10 mR/hr	9.2 mR/hr	10.1 mR/hr
Rate	4 mR/hr	3.2 mR/hr	3.6 mR/hr
Rate	1 mR/hr	0.9 mR/hr	0.95 mR/hr
Rate	400 µR/hr	370 µR/hr	370 µR/hr
Rate	100 µR/hr	103 µR/hr	103 µR/hr
Integrate	80 mR	69 mR	81 mR
Integrate	20 mR	16.6 mR	19 mR


Readings with * indicate ranges where As-Found readings are >20% of Target value. Readings with ** indicate As-left readings are >10.00% of Target value

Outer Physical Check: <i>Pass</i>	Tap Test: <i>Pass</i>
Internal Check: <i>Pass</i>	


Comments: All readings 10 mR/hr or higher and the integrate readings were obtained using the J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N: 9141). All other readings were obtained using the Tech Ops Model 773 Cs-137 Beam Calibrator (S/N: S-1110).

As-Found/As-Left Calibration Factors: F1 = 107/107, F2 = 115/120, F3 = 110/115, F4 = 100/117, F5 = 92/105, INT = 100/100

Chris Pirie
Calibration
Technician



QA
Review:



Calibration Date: 04/10/2019
Expires: 04/10/2020

Atmospheric Conditions - Temperature: 71°F Humidity: 33% Barometric Pressure: 29.64"hg

This calibration was performed by RSCS using one or more of the following NIST Traceable radiation sources:

Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0%.

J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7%.

RSCS Neutron Calibrator, AmBe Source Model NUMEC-AM-31 (S/N Am-478), characterized using Far West Technologies Model FWAD-1 "HAWK" TEPC (S/N 021) in accordance with the methods specified in RSCS TSD 13-002, with estimated uncertainty of 9.4%.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R.

RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2005 while demonstrating technical competence in the field of calibration. Refer to the Scope of Accreditation AC-2079 for information on the types of calibrations to which this accreditation applies.

This calibration certificate shall not be reproduced except in full without the express written consent of RSCS, Inc.

Customer: Jen Newtown
Upstate Medical Physics
1290 Blossom Drive
Victor, NY 14564

Instrument
Victoreen Model 451P

Serial Number
3869

Precision Check				
Test 1	Test 2	Test 3	Mean	Results
380.00 mR/hr	380.00 mR/hr	380.00 mR/hr	380.00 mR/hr	Satisfactory

Accuracy Check			
Range	Target Value	As Found	As Left
Rate	4 R/hr	3.5 R/hr	4.1 R/hr
Rate	1 R/hr	.87 R/hr	1.00 R/hr
Rate	400 mR/hr	330 mR/hr	380 mR/hr
Rate	100 mR/hr	80 mR/hr	100 mR/hr
Rate	40 mR/hr	40 mR/hr	40 mR/hr
Rate	10 mR/hr	10 mR/hr	10 mR/hr
Rate	4 mR/hr	4 mR/hr	4 mR/hr
Rate	1 mR/hr	1.03 mR/hr	1.03 mR/hr
Rate	400 µR/hr	400 µR/hr	400 µR/hr
Rate	100 µR/hr	102 µR/hr	102 µR/hr
Integrate	80 mR	72 mR	83 mR
Integrate	20 mR	17.6 mR	19.6 mR

Readings with * indicate ranges where As-Found readings are >20% of Target value. Readings with ** indicate As-left readings are >10.00% of Target value

Outer Physical Check: Pass

Tap Test: Pass


Comments: All readings higher than 40 mR/hr were obtained using the J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N: 9141). All other readings were obtained using the Tech Ops Model 773 Cs-137 Beam Calibrator (S/N: S-1110).

As-Found/As-Left Calibration Factors: F1 = 95/95, F2 = 111/111, F3 = 106/106, F4 = 80/90, F5 = 80/90, INT = 100/100

Chris Hartman
Calibration
Technician



QA
Review:



Calibration Date: 05/18/2023
Expires: 05/18/2024

Atmospheric Conditions - Temperature: 68°F Humidity: 35% Barometric Pressure: 30.16"hg

This calibration was performed by RSCS using one or more of the following NIST Traceable radiation sources:
Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0%.
J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7%.
J.L. Shepherd and Associates Model 81-12 Cs-137 Dual Source Gamma Calibrator (S/N 7145), characterized using Exradin Models A6 (S/N 185) / A4 (S/N 220) / A5 (S/N 261), GE RSS-131 (S/N 95100178), and Standard Imaging Electrometer Model Maxx4000 (S/N J171510) in accordance with methods specified in RSCS TSD 20-023, with estimated uncertainty of 6.7%.
RSCS Model 149 Neutron Calibrator, Pu:Be Source Model MRC Model 2726-B (S/N 495), characterized using Far West Technologies Model Rem500 TEPC (S/N 271) in accordance with the methods specified in RSCS TSD 21-049, with an uncertainty of 9.4% at a 95% confidence interval.
The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement.
Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R.
RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.
Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994
RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2005 while demonstrating technical competence in the field of calibration. Refer to the Scope of Accreditation AC-2079 for information on the types of calibrations to which this accreditation applies.
This calibration certificate shall not be reproduced except in full without the express written consent of RSCS, Inc.

MAY 18 2023

Customer: Jen Newtown
Upstate Medical Physics
1290 Blossom Drive
Victor NY 14564

Instrument
Victoreen Model 451B

Serial Number
799

Precision Check				
Test1	Test2	Test3	Mean	Results
41.00 mR/hr	41.00 mR/hr	41.00 mR/hr	41.00 mR/hr	Pass

Accuracy Check				
Range	Target Value	As Found	As Left	Source ID
Rate	40 R/hr	38 R/hr	37 R/hr	R-028/R-122
Rate	10 R/hr	11.1 R/hr	10.8 R/hr	R-028/R-122
Rate	4 R/hr	4 R/hr	4 R/hr	R-028/R-122
Rate	1 R/hr	1.05 R/hr	1.04 R/hr	R-028/R-122
Rate	400 mR/hr	410 mR/hr	400 mR/hr	R-028/R-122
Rate	100 mR/hr	109 mR/hr	107 mR/hr	R-028/R-122
Rate	40 mR/hr	39 mR/hr	41 mR/hr	R-137
Rate	10 mR/hr	9.8 mR/hr	10.4 mR/hr	R-137
Rate	4 mR/hr	4 mR/hr	4 mR/hr	R-137
Rate	1 mR/hr	1.04 mR/hr	1.04 mR/hr	R-137
Integrate	200 mR	200 mR	200 mR	R-028/R-122
Integrate	50 mR	52 mR	52 mR	R-028/R-122

Readings with * indicate ranges where As-Found readings are >20% of Target value.

Readings with ** indicate As-left readings are >10% of Target value.

Readings with # indicate ranges were calibrated using a pulser

Physical Checks
Outer Physical Check: Pass
Internal Physical Check: Pass
Tap Test: Pass
Desiccant Check: Pass

Comments: As-Found/As-Left Calibration Factors: F1 = 100/100, F2 = 90/96, F3 = 102/99, F4 = 95/95, F5 = 95/93, INT = (100/100)/(100/100). Dried / Replaced Desiccant.

Customer Equipment ID: 00494

Calibrated By Chris Hartman



QC Review By Matthew Hunter



Date: Jan 23, 2024

Expires: Jan 23, 2025

QC Date: Jan 23, 2024

Atmospheric Conditions - Temperature: 68.4 °F, Humidity: 19.0 %, Barometric Pressure: 30.23 in/hg.

In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. This calibration was performed by RSCS using the following NIST Traceable sources:

R-028/R-122 - J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7% (Cal Due: Feb 02, 2024)

R-137 - Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0% (Cal Due: Feb 02, 2024)

Worked on calibration: Chris Hartman

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCCL Z540-1-1994.

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2017 while demonstrating technical competence in the field of

Customer:
Upstate Medical Physics
1290 Blossom Drive
Victor, NY 14564**Instrument**
Ludlum Model 9DP-1**Serial Number**
25006598

Precision Check				
Test 1	Test 2	Test 3	Mean	Results
38.70 mR/hr	38.50 mR/hr	38.70 mR/hr	38.63 mR/hr	Satisfactory

Accuracy Check			
Range	Target Value	As Found	As Left
X10K	40 R/hr	40.4 R/hr	40.4 R/hr
X10K	10 R/hr	10.9 R/hr	10.9 R/hr
X1K	4 R/hr	4.29 R/hr	4.29 R/hr
X1K	1 R/hr	1.10 R/hr	1.10 R/hr
X100	400 mR/hr	428 mR/hr	428 mR/hr
X100	100 mR/hr	103 mR/hr	103 mR/hr
X10	40 mR/hr	38.7 mR/hr	38.7 mR/hr
X10	10 mR/hr	10.5 mR/hr	10.5 mR/hr
X1	4 mR/hr	3.91 mR/hr	4.32 mR/hr
X1	1 mR/hr	0.86 mR/hr	0.91 mR/hr

Readings with * indicate ranges where As-Found readings are >20% of Target value. Readings with ** indicate As-left readings are >10.00% of Target value

Outer Physical Check: Pass

Tap Test: Pass

Comments: All readings 10 mR/hr or higher were obtained using the J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N: 9141). All other readings were obtained using the Tech Ops Model 773 Cs-137 Beam Calibrator (S/N: S-1110).

As Found/As Left Calibration Factor: X1 = 806/900 X10 = 810/810 X100 = 857/857 X1K = 888/888 X10K = 1367/1367

Mark Nelson
Calibration
Technician



QA
Review:



Calibration Date: 01/31/2020
Expires: 01/31/2021

Atmospheric Conditions - Temperature: 72°F Humidity: 26% Barometric Pressure: 30.22"hg

This calibration was performed by RSCS using one or more of the following NIST Traceable radiation sources:

Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0%.

J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7%.

RSCS Neutron Calibrator, AmBe Source Model NUMEC-AM-31 (S/N Am-478), characterized using Far West Technologies Model FWAD-1 "HAWK" TEPC (S/N 021) in accordance with the methods specified in RSCS TSD 13-002, with estimated uncertainty of 9.4%

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2005 while demonstrating technical competence in the field of calibration. Refer to the Scope of Accreditation AC-2079 for information on the types of calibrations to which this accreditation applies.

This calibration certificate shall not be reproduced except in full without the express written consent of RSCS, Inc.



www.ludlums.com

Designer and Manufacturer
of
Scientific and Industrial
Instruments

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.

 501 Oak Street
325-235-5494

Sweetwater, TX 79556, U.S.A.



CERT # 4084.01

Customer LANDAUER MEDICAL PHYSICSORDER NO. 20440721/536615Mfg. Fluke Model 451P-RYR Serial No. 7206

Mfg. _____ Model _____ Serial No. _____

Cal. Date 18-Feb-23 Cal Due Date 18-Feb-24 Cal. Interval 1 Year Meterface DigitalCheck mark ☒ applies to applicable instr. and/or detector IAW mfg. spec. T. 76 °F RH 15 % Alt 706.6 mm Hg
☐ New Instrument ☐ Instrument Received ☐ Within Toler. +/-10% ☐ 10-20% ☐ Out of Tol. ☐ Requiring Repair ☒ Other-See comments

☒ Mechanical ck. ☐ Meter Zeroed ☐ Background Subtract ☐ Input Sens. Linearity

☐ F/S Resp. ck. ☐ Reset ck. ☐ Window Operation ☐ Geotropism

☐ Audio ck. ☐ Alarm Setting ck. ☐ Batt. ck.
☐ Calibrated in accordance with LMI SOP 14.8☒ Calibrated in accordance with LMI SOP 14.9

Instrument Volt Set _____ V Input Sens. _____ mV Det. Oper. _____ V at _____ mV Threshold Dial Ratio _____ = _____ mV

☐ HV Readout (2 points) Ref./Inst. _____ / _____ V Ref./Inst. _____ / _____ V

COMMENTS:

Instrument received with chamber depressurized.

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

Multimeter uncertainty within 1.3% of reading, Gamma uncertainty within 5.0% of reading, Neutron uncertainty within 7.0% of reading, Count rate uncertainty within 5.4% of reading

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			
Digital			

Range(s) Calibrated Electronically

Digital	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING	Log	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING
Readout	2 R/hr	<u>NA</u>	<u>2.0 R/hr</u>	Scale			
	200 mR/hr		<u>200 mR/hr</u>				
	20 mR/hr		<u>20</u>				
	2 mR/hr		<u>2.0</u>				
	200 µR/hr		<u>200 µR/hr</u>				

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques.

All pass/fail determinations are based on the manufacturer's specifications without considering uncertainty factors.
Measurement results represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k=2.

 ISO/IEC 17025:2017(E)
State of Texas Calibration License No. LO-1963

The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323AB-2013

Reference Instruments and/or Sources: Cs-137 S/N: ☐ 059 ☐ 2171CP ☐ 2261CP ☐ 720 ☐ 734 ☐ 781 ☐ 1131 ☐ 1616 ☐ 1696 ☐ 1909 ☐ 1916CP ☒ 2324/2521
☐ 5717CO ☐ 5719CO ☐ 60646 ☐ 70897 ☐ 73410 ☐ E552 ☒ G112 ☒ 2168CP ☐ S-394 ☐ S-1054 ☐ T10081 ☐ T10082 Neutron Am-241 Be ☐ T-304 Ra-226 ☐ Y982
☐ E551 ☐ 5105 ☐ CSV280

☐ Alpha S/N _____ ☐ Beta S/N _____ ☐ Other _____

☐ m 500 S/N _____ ☐ Oscilloscope S/N _____ ☐ Multimeter S/N _____
Calibrator Donnie Miekos Title Calibrator Date 18 Feb 23QC'd By Rid. 6 Title Final QC Date 20 Feb 23

Customer: Jen Newtown
Upstate Medical Physics
1290 Blossom Drive
Victor NY 14564

Instrument
Ludlum Model 3-241R

Serial Number
264306

Probe Model

Ludlum 44-9

Ludlum 44-38

Serial Number

388694

567

Precision Check				
Test1	Test2	Test3	Mean	Results
4.00 mR/hr	4.00 mR/hr	4.00 mR/hr	4.00 mR/hr	Pass

Accuracy Check				
Range	Target Value	As Found	As Left	Source ID
X100	160 mR/hr	175 mR/hr	175 mR/hr	R-200/R-201
X100	40 mR/hr	40 mR/hr	40 mR/hr	R-200/R-201
X10	16 mR/hr	16.5 mR/hr	16.5 mR/hr	R-200/R-201
X10	4 mR/hr	4 mR/hr	4 mR/hr	R-200/R-201
X1	1.6 mR/hr	1.75 mR/hr	1.75 mR/hr	R-200/R-201
X1	0.4 mR/hr	0.4 mR/hr	0.4 mR/hr	R-200/R-201
X.1	0.16 mR/hr #	0.19 mR/hr #	0.16 mR/hr #	Pulser 98756
X.1	0.04 mR/hr #	0.045 mR/hr #	0.04 mR/hr #	Pulser 98756

Readings with * indicate ranges where As-Found readings are >20% of Target value.

Readings with ** indicate As-left readings are >10% of Target value.

Readings with # indicate ranges were calibrated using a pulser

Probe	Isotope	Efficiency	NIST Source ID	Geometry	Comment
44-9 388694	Tc-99m	0.0020 C/D (4 Pi)	1867-29-1 Co-57	1 cm	

Physical Checks
Outer Physical Check: Pass
Internal Physical Check: Pass
Geotropism Check: Pass
Mechanical Zero Check: Pass
Tap Test: Pass
Audio Function Check: Pass
Cord check: Pass

Electronics Checks	As Found	As Left
High Voltage	901 Volts	901 Volts


Comments: 44-9 Probe is NOT calibrated in mR/hour!

Customer Equipment ID:

Calibrated By Chris Pirie



QC Review By Matt
Toleos



Date: Feb 21, 2024

Expires: Feb 21, 2025

QC Date: Feb 21, 2024

Atmospheric Conditions - Temperature: 70.2 °F, Humidity: 20.0 %, Barometric Pressure: 30.47 in/hg.

In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement.

This calibration was performed by RSCS using the following NIST Traceable sources:

Pulser 98756 - (Cal Due: Oct 31, 2024)

R-200/R-201 - J.L. Shepherd and Associates Model 81-12 Cs-137 Dual Source Gamma Calibrator (SN 7145), characterized using Exradin Models A6 (S/N 185) / A4 (S/N 220) / A5 (S/N 261), GE RSS-131 (S/N 95100178), and Standard Imaging Electrometer Model Maxx4000 (S/N J171510) in accordance with methods specified in RSCS TSD 20-059, with estimated uncertainty of 6.26%. (Cal Due: Jan 12, 2025)

Worked on calibration: Chris Pirie

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994.

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2017 while demonstrating technical competence in the field of calibration. Refer to the Scope of Accreditation AC-2079 for information on the types of calibrations to which this accreditation applies.

This calibration certificate shall not be reproduced except in full without the express written consent of RSCS, Inc.