

Certificate number: o60252997i232730-1



# Calibration Certificate

Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 Light  
Serial number 232730  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 5/29/2025  
As left 5/29/2025  
Adjustment done Yes  
Tested by Amber Ramos

Approved by



Kierra Sullivan  
Lab Technician

Certificate date 6/9/2025

## CALIBRATION RESULT

AS FOUND: One or more measured and tested values of this certificate are within the specifications by a margin less than the expanded measurement uncertainty. Therefore, it is not possible to state conformance with specifications for those measurements.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Illuminance: 5 %  
Luminance: 5 %



Certificate number: o60253110i230095-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 Survey  
Serial number 230095  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 6/2/2025  
As left 6/2/2025  
Adjustment done Yes  
Tested by Amber Ramos

Approved by

Kierra Sullivan  
Lab Technician

Certificate date 6/9/2025

## CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60253083i263739-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 CT  
Serial number 263739  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 6/2/2025  
As left 6/2/2025  
Adjustment done Yes  
Tested by Amber Ramos

Approved by

Kierra Sullivan  
Lab Technician

Certificate date 6/9/2025

## CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60253009i267808-B-1



# Calibration Certificate

Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 267808  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 5/29/2025  
As left 5/30/2025  
Adjustment done Yes  
Tested by ~~Armud Bryant~~  
Christina Arroyo

Approved by



Kierra Sullivan  
Lab Technician  
6/9/2025

Certificate date



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com


## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 267808  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 5/29/2025  
As left 5/30/2025  
Adjustment done Yes  
Tested by Armad Bryant  
Christina Arroyo

Approved by



Kierra Sullivan  
Lab Technician

Certificate date 6/9/2025

## CALIBRATION RESULT

- AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.
- AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

### Relevant instrument specifications

Air Kerma: 5 %  
Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60253024i246291-B-1



# Calibration Certificate

Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 R/F  
Serial number 246291  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 5/29/2025  
As left 5/30/2025  
Adjustment done Yes  
Tested by Amber Ramos  
Armada Bryant

Approved by

A blue ink signature of Kierra Sullivan, written in a cursive style, is positioned above a horizontal line.

Kierra Sullivan  
Lab Technician

Certificate date 6/9/2025



Certificate number: o60253024i246291-A-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 R/F  
Serial number 246291  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 5/29/2025  
As left 5/30/2025  
Adjustment done Yes  
Tested by Amber Ramos  
Armada Bryant  
  
Approved by   
Kierra Sullivan  
Lab Technician  
  
Certificate date 6/9/2025

## CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.  
AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications  
Air kerma: 5 %, Air kerma rate: 5 %  
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

