

Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DRIVE 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 335155
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2025-04-04
Adjustment done Yes
Tested by Vera Zhang

Approved by

Rasmus Brännström

Rasmus Brännström
Finalization operator

Certificate date

2025-05-05

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DRIVE 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 335155
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2025-04-04
Adjustment done Yes
Tested by Vera Zhang

Approved by

Rasmus Brännström

Rasmus Brännström
Finalization operator

Certificate date 2025-05-05

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60252040i279408-B-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 279408
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 5/5/2025
As left 5/5/2025
Adjustment done Yes
Tested by Amber Ramos
Armada Bryant
Christina Arroyo

Approved by



Kierra Sullivan

Lab Technician

Certificate date 5/7/2025



Certificate number: o60252040i279408-A-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 279408
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 5/5/2025
As left 5/5/2025
Adjustment done Yes
Tested by Amber Ramos
Armad Bryant
Christina Arroyo

Approved by

Kierra Sullivan
Lab Technician

Certificate date 5/7/2025

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air Kerma: 5 %

Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60251830i263381-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 CT
Serial number 263381
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 4/29/2025
As left 4/29/2025
Adjustment done Yes
Tested by Armad Bryant

Approved by

Kierra Sullivan
Lab Technician

Certificate date 5/7/2025

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DRIVE 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Light
Serial number 325746
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2025-04-07
Adjustment done Yes
Tested by Frida Lindgren

Approved by

Rasmus Brännström

Rasmus Brännström
Finalization operator

Certificate date

2025-05-05

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Illuminance: 5 %
Luminance: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DRIVE 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Survey
Serial number 331795
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2025-03-27
Adjustment done Yes
Tested by Ida Wiholm

Approved by

Rasmus Brännström

Rasmus Brännström
Finalization operator

Certificate date 2025-05-05

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

