

Certificate number: o60254321i273477-A-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 273477
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 7/9/2025
As left 7/9/2025
Adjustment done Yes
Tested by Amber Ramos

Approved by

Josefina Diaz
Finalization technician

Certificate date 7/22/2025

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.
AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60254321i273477-B-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 273477
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 7/9/2025
As left 7/9/2025
Adjustment done Yes
Tested by Amber Ramos

Approved by


Josefine Diaz
Finalization technician

Certificate date 7/22/2025



Certificate number: o60254410i268286-A-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 268286
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 7/11/2025
As left 7/11/2025
Adjustment done Yes
Tested by Tommy Cox

Approved by


Josefina Diaz
Finalization technician

Certificate date 7/22/2025

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air Kerma: 5 %

Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60254410i268286-B-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 268286
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 7/11/2025
As left 7/11/2025
Adjustment done Yes
Tested by Tommy Cox

Approved by


Josefine Diaz
Finalization technician

Certificate date 7/22/2025



Certificate number: o60254246i269921-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 CT
Serial number 269921
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 7/8/2025
As left 7/8/2025
Adjustment done Yes
Tested by Maria Isabel Magana Guerrero

Approved by


Josefina Diaz
Finalization technician

Certificate date

7/22/2025

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60254346i272201-1



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Light
Serial number 272201
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 7/10/2025
As left 7/11/2025
Adjustment done Yes
Tested by Christina Arroyo

Approved by


Josefina Diaz
Finalization technician

Certificate date 7/22/2025

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Illuminance: 5 %

Luminance: 5 %



Certificate number: o60254158i262620-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe Inc.
2 Science Road
Glenwood, IL 60425-1586
USA
+1-833-296-9240
customerservice.us@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Survey
Serial number 262620
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 7/7/2025
As left 7/7/2025
Adjustment done Yes
Tested by Amber Ramos

Approved by


Josefina Diaz
Finalization technician

Certificate date 7/22/2025

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

