

Certificate number: o60257053i273430-B-1



# Calibration Certificate

Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 R/F  
Serial number 273430  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 9/19/2025  
As left 9/19/2025  
Adjustment done Yes  
Tested by Nate Barnes

Approved by

A handwritten signature in blue ink, appearing to read "Afsana Arbab", is written over a horizontal line.

Afsana Arbab  
Service Technician

Certificate date 10/20/2025



Certificate number: o60257053i273430-A-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 R/F  
Serial number 273430  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 9/19/2025  
As left 9/19/2025  
Adjustment done Yes  
Tested by Nate Barnes

Approved by

  
Afsana Arbab  
Service Technician

Certificate date 10/20/2025

## CALIBRATION RESULT

**AS FOUND:** All measured and tested values of this certificate were found to be in conformance with the specification.

**AS LEFT:** All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %  
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe AB  
Björklundabacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 268368  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2025-10-07  
As left 2025-10-11  
Adjustment done Yes  
Tested by Hyeji Hwang  
Nils Liljegren

Approved by



Afsana Arbab  
Service Technician

Certificate date 2025-10-20

## CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air Kerma: 5 %  
Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe AB  
Björklundabacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 268368  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2025-10-07  
As left 2025-10-11  
Adjustment done Yes  
Tested by Hyeji Hwang  
Nils Liljegren

Approved by



Afsana Arbab  
Service Technician

Certificate date

2025-10-20

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60256845i273383-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 CT  
Serial number 273383  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 9/16/2025  
As left 9/16/2025  
Adjustment done Yes  
Tested by Armad Bryant

Approved by

Afsana Arbab  
Service Technician

Certificate date 10/20/2025

## CALIBRATION RESULT

**AS FOUND:** All measured and tested values of this certificate were found to be in conformance with the specification.

**AS LEFT:** All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o60256909i272340-1



# Calibration Certificate

Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US


Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 Light  
Serial number 272340  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 9/17/2025  
As left 9/17/2025  
Adjustment done Yes  
Tested by Amber Ramos

Approved by   
Afsana Arbab  
Service Technician

Certificate date 10/20/2025

## CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Illuminance: 5 %

Luminance: 5 %



Certificate number: o60256836i262957-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
BLOSSOM DR 1290  
14564 VICTOR  
US

Laboratory  
Unfors RaySafe Inc.  
2 Science Road  
Glenwood, IL 60425-1586  
USA  
+1-833-296-9240  
customerservice.us@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 Survey  
Serial number 262957  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 9/16/2025  
As left 9/16/2025  
Adjustment done Yes  
Tested by Armad Bryant

Approved by

Afsana Arbab  
Service Technician

Certificate date 10/20/2025

## CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the American Association for Laboratory Accreditation (A2LA) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

