

Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US


Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 mAs
Serial number 307805
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2025-10-29
As left 2025-11-01
Adjustment done Yes
Tested by Emmy Pantzar
Philip Kinnander

Approved by 

Frida Lindgren
Lab Technician

Certificate date 2025-11-04

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Electrical charge: 1 %
Electrical current: 1 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 331305
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2025-11-02
As left 2025-11-02
Adjustment done Yes
Tested by Anton Johansson

Approved by



Frida Lindgren
Lab Technician

Certificate date 2025-11-04

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 331305
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2025-11-02
As left 2025-11-02
Adjustment done Yes
Tested by Anton Johansson

Approved by



Frida Lindgren
Lab Technician

Certificate date 2025-11-04

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 268452
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2025-11-02
As left 2025-11-04
Adjustment done Yes
Tested by Kristoffer Petric
Nils Liljegen

Approved by



Marie Panula
Service Technician

Certificate date

2025-11-05

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air Kerma: 5 %
Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 268452
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2025-11-02
As left 2025-11-04
Adjustment done Yes
Tested by Kristoffer Petric
Nils Liljegren

Approved by



Marie Panula
Service Technician

Certificate date

2025-11-05

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 CT
Serial number 273843
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2025-10-29
As left 2025-11-01
Adjustment done Yes
Tested by Ujjaini Joseph

Approved by 
Frida Lindgren
Lab Technician
Certificate date 2025-11-04

CALIBRATION RESULT

- AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.
- AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications
Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Light
Serial number 272096
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2025-10-29
As left 2025-10-31
Adjustment done Yes
Tested by Emmy Pantzar
Philip Kinnander

Approved by



Frida Lindgren
Lab Technician

Certificate date 2025-11-04

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Illuminance: 5 %
Luminance: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DR 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Survey
Serial number 262935
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2025-10-29
As left 2025-10-29
Adjustment done Yes
Tested by Hussein Mahamod

Approved by



Frida Lindgren
Lab Technician

Certificate date 2025-11-04

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

