

Certificate number: o70056279i275762-B-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14565 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 275762
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-15
As left 2026-01-27
Adjustment done Yes
Tested by Erik Strömberg

Approved by

Frida Lindgren
Lab Technician

Certificate date

2026-01-28

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o70056279i275762-A-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14565 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 275762
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-15
As left 2026-01-27
Adjustment done Yes
Tested by Erik Strömberg

Approved by

Frida Lindgren
Lab Technician

Certificate date 2026-01-28

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o70057338i273279-B-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14565 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 273279
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-25
As left 2026-01-25
Adjustment done Yes
Tested by Ujjaini Joseph
Vera Zhang

Approved by

Frida Lindgren
Lab Technician

Certificate date

2026-01-28

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o70057338i273279-A-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14565 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 273279
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-25
As left 2026-01-25
Adjustment done Yes
Tested by Ujjaini Joseph
Vera Zhang

Approved by

Frida Lindgren
Lab Technician

Certificate date

2026-01-28

CALIBRATION RESULT

AS FOUND: One or more measured and tested values of this certificate are within the specifications by a margin less than the expanded measurement uncertainty. Therefore, it is not possible to state conformance with specifications for those measurements.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air Kerma: 5 %

Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14565 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 CT
Serial number 273976
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-10
As left 2026-01-27
Adjustment done Yes
Tested by Erik Strömberg
Ujjaini Joseph

Approved by



Frida Lindgren
Lab Technician

Certificate date 2026-01-28

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o70056931i272035-1



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14565 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Light
Serial number 272035
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-21
As left 2026-01-26
Adjustment done Yes
Tested by Emmy Pantzar

Approved by

Frida Lindgren
Lab Technician

Certificate date

2026-01-28

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Illuminance: 5 %
Luminance: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14565 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Survey
Serial number 262942
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-09
As left 2026-01-26
Adjustment done Yes
Tested by Hussein Mahamod
Ujjaini Joseph

Approved by



Frida Lindgren
Lab Technician

Certificate date 2026-01-28

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

