



Calibration Certificate - ID Number: 26430600000 - 20260130 - 0

Customer: Jen Schmitt
 Upstate Medical Physics
 6536 Anthony Dr
 Suite B
 Victor NY 14564

Instrument	Serial Number
Ludlum Model 3-241R	264306
Probe Model	Serial Number
Ludlum 44-38	567
Ludlum 44-9	388694

Precision Check				
Test1	Test2	Test3	Mean	Results
40.00 mR/hr	40.00 mR/hr	40.00 mR/hr	40.00 mR/hr	Pass

Accuracy Check				
Range	Target Value	As Found	As Left	Source ID
X100	160 mR/hr	175 mR/hr	175 mR/hr	R-137
X100	40 mR/hr	40 mR/hr	40 mR/hr	R-137
X10	16 mR/hr	16 mR/hr	16 mR/hr	R-137
X10	4 mR/hr	4 mR/hr	4 mR/hr	R-137
X1	1.6 mR/hr	1.55 mR/hr	1.55 mR/hr	R-137
X1	0.4 mR/hr	0.4 mR/hr	0.4 mR/hr	R-137
X.1	0.16 mR/hr #	0.15 mR/hr #	0.17 mR/hr #	4084.01
X.1	0.04 mR/hr #	0.035 mR/hr #	0.04 mR/hr #	4084.01

Readings with * indicate ranges where As-Found readings are >20% of Target value.
 Readings with ** indicate As-left readings are >10% of Target value.
 Readings with # indicate ranges were calibrated using a pulser

Probe	Isotope	Efficiency	NIST Source ID	Geometry	Comment
44-9 388694	Tc-99m	0.0003 C/D (4 Pi)	2501-1 Co-57	1 cm	

Physical Checks
Outer Physical Check: Pass
Internal Physical Check: Pass
Geotropism Check: Pass
Mechanical Zero Check: Pass
Tap Test: Pass
Audio Function Check: Pass
Cable Check: Pass

Electronics Checks	As Found	As Left
High Voltage	900 Volts	900 Volts

Comments:
 44-38 probe calibrated perpendicular to the source with window closed
 Customer Equipment ID:

Calibrated By Ned Howard

QC Review By Jarod Fagan

Date: Jan 30, 2026
 Expires: Jan 30, 2027
 QC Date: Feb 02, 2026

Atmospheric Conditions - Temperature: 72.5 °F, Humidity: 19.8 %, Barometric Pressure: 29.85 in/hg.
 In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated.
 Unless otherwise stated, this calibration was performed by RSCS at 93 Ledge Rd, Seabrook, NH utilizing the following NIST Traceable sources:
 4084.01 - (Cal Due: May 22, 2026)
 R-137 - Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0% (Cal Due: Aug 29, 2026)

Worked on calibration: Ned Howard
 The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.





Customer: Jen Schmitt
 Upstate Medical Physics
 1290 Blossom Drive
 Victor NY 14564

Instrument
 Victoreen Model 451B

Serial Number
 452

Precision Check				
Test1	Test2	Test3	Mean	Results
1.05 R/hr	1.05 R/hr	1.05 R/hr	1.05 R/hr	Pass

Accuracy Check				
Range	Target Value	As Found	As Left	Source ID
Rate	40 R/hr	36 R/hr	36 R/hr	R-028/R-122
Rate	10 R/hr	10.3 R/hr	10.3 R/hr	R-028/R-122
Rate	4 R/hr	4.1 R/hr	4.1 R/hr	R-028/R-122
Rate	1 R/hr	1.05 R/hr	1.05 R/hr	R-028/R-122
Rate	400 mR/hr	390 mR/hr	390 mR/hr	R-028/R-122
Rate	100 mR/hr	106 mR/hr	106 mR/hr	R-028/R-122
Rate	40 mR/hr	38 mR/hr	38 mR/hr	R-137
Rate	10 mR/hr	9.8 mR/hr	9.8 mR/hr	R-137
Rate	4 mR/hr	3.7 mR/hr	3.7 mR/hr	R-137
Rate	1 mR/hr	0.95 mR/hr	0.95 mR/hr	R-137
Integrate	200 mR	200 mR	200 mR	R-028/R-122
Integrate	50 mR	52 mR	52 mR	R-028/R-122

Readings with * indicate ranges where As-Found readings are >20% of Target value.

Readings with ** indicate As-left readings are >10% of Target value.

Readings with # indicate ranges were calibrated using a pulser

Physical Checks
Outer Physical Check: Pass
Internal Physical Check: Pass
Tap Test: Pass
Desiccant Check: Pass

Comments:

Customer Equipment ID:

Calibrated By Ned Howard

QC Review By Jarod Fagan

Date: Oct 01, 2025

Expires: Oct 01, 2026

QC Date: Oct 01, 2025

Atmospheric Conditions - Temperature: 70.0 °F, Humidity: 53.4 %, Barometric Pressure:30.11 in/hg.

In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated.

Unless otherwise stated, this calibration was performed by RSCS at 93 Ledge Rd, Seabrook, NH utilizing the following NIST Traceable sources: R-028/R-122 - J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7% (Cal Due: Jan 03, 2026)

R-137 - Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0% (Cal Due: Jan 03, 2026)

Worked on calibration: Ned Howard

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994.

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2017 while demonstrating technical competence in the field of



Customer: Jen Schmitt
Upstate Medical Physics
1290 Blossom Drive
Victor NY 14564

Instrument
Victoreen Model 451B

Serial Number
1074

Precision Check				
Test1	Test2	Test3	Mean	Results
103.00 mR/hr	103.00 mR/hr	103.00 mR/hr	103.00 mR/hr	Pass

Accuracy Check				
Range	Target Value	As Found	As Left	Source ID
Rate	40 R/hr	36 R/hr	36 R/hr	R-028/R-122
Rate	10 R/hr	10.7 R/hr	10.7 R/hr	R-028/R-122
Rate	4 R/hr	3.8 R/hr	3.8 R/hr	R-028/R-122
Rate	1 R/hr	1.06 R/hr	1.06 R/hr	R-028/R-122
Rate	400 mR/hr	370 mR/hr	370 mR/hr	R-028/R-122
Rate	100 mR/hr	103 mR/hr	103 mR/hr	R-028/R-122
Rate	40 mR/hr	39 mR/hr	39 mR/hr	R-137
Rate	10 mR/hr	10.2 mR/hr	10.2 mR/hr	R-137
Rate	4 mR/hr	4 mR/hr	4 mR/hr	R-137
Rate	1 mR/hr	1.05 mR/hr	1.05 mR/hr	R-137
Integrate	200 mR	200 mR	200 mR	R-028/R-122
Integrate	50 mR	52 mR	52 mR	R-028/R-122

Readings with * indicate ranges where As-Found readings are >20% of Target value.
Readings with ** indicate As-left readings are >10% of Target value.
Readings with # indicate ranges were calibrated using a pulser

Physical Checks
Outer Physical Check: Pass
Internal Physical Check: Pass
Tap Test: Pass
Desiccant Check: Pass (Replaced.)

Comments:

Customer Equipment ID:

Calibrated By Jarod Fagan

QC Review By Matthew Hunter

Date: Jul 10, 2025

Expires: Jul 10, 2026

QC Date: Jul 10, 2025

Atmospheric Conditions - Temperature: 71.9 °F, Humidity: 52.5 %, Barometric Pressure:29.90 in/hg.

In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated.

Unless otherwise stated, this calibration was performed by RSCS at 93 Ledge Rd, Seabrook, NH utilizing the following NIST Traceable sources: R-028/R-122 - J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7% (Cal Due: Jan 03, 2026)

R-137 - Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0% (Cal Due: Jan 03, 2026)

Worked on calibration: Jarod Fagan

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994.

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2017 while demonstrating technical competence in the field of



Calibration Certificate - ID Number: 38690000000 - 20250527 - 0

Customer: Jen Schmitt
Upstate Medical Physics
1290 Blossom Drive
Victor NY 14564

Instrument
Victoreen Model 451P

Serial Number
3869

Table with 5 columns: Test1, Test2, Test3, Mean, Results. Row 1: 3.80 mR/hr, 3.70 mR/hr, 3.80 mR/hr, 3.77 mR/hr, Pass.

Table with 5 columns: Range, Target Value, As Found, As Left, Source ID. Multiple rows listing various rate and integrate tests and their results.

Readings with * indicate ranges where As-Found readings are >20% of Target value.
Readings with ** indicate As-left readings are >10% of Target value.
Readings with # indicate ranges were calibrated using a pulser

Table with 1 column: Physical Checks. Rows: Outer Physical Check: Pass, Tap Test: Pass.

Comments:

Customer Equipment ID:

Calibrated By Jarod Fagan

[Signature]

QC Review By Matthew Hunter

[Signature]

Date: May 27, 2025

Expires: May 27, 2026

QC Date: May 27, 2025

Atmospheric Conditions - Temperature: 71.0 °F, Humidity: 47.5 %, Barometric Pressure:30.20 in/hg.

In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated.

Unless otherwise stated, this calibration was performed by RSCS at 93 Ledge Rd, Seabrook, NH utilizing the following NIST Traceable sources: R-028/R-122 - J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7% (Cal Due: Jan 03, 2026)

R-137 - Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0% (Cal Due: Jan 03, 2026)

Worked on calibration: Jarod Fagan

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994.

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2017 while demonstrating technical competence in the field of calibration. Refer to the Scope of Accreditation AC-2079 for information on the types of calibrations to which this accreditation applies.

This calibration certificate shall not be reproduced except in full without the express written consent of RSCS, Inc.



Customer: Jen Newtown
 Upstate Medical Physics
 1290 Blossom Drive
 Victor NY 14564

Instrument
 Victoreen Model 451B

Serial Number
 799

Precision Check				
Test1	Test2	Test3	Mean	Results
10.30 mR/hr	10.30 mR/hr	10.30 mR/hr	10.30 mR/hr	Pass

Accuracy Check				
Range	Target Value	As Found	As Left	Source ID
Rate	40 R/hr	37 R/hr	37 R/hr	R-028/R-122
Rate	10 R/hr	10.7 R/hr	10.7 R/hr	R-028/R-122
Rate	4 R/hr	4 R/hr	4 R/hr	R-028/R-122
Rate	1 R/hr	1.05 R/hr	1.05 R/hr	R-028/R-122
Rate	400 mR/hr	400 mR/hr	400 mR/hr	R-028/R-122
Rate	100 mR/hr	106 mR/hr	106 mR/hr	R-028/R-122
Rate	40 mR/hr	40 mR/hr	40 mR/hr	R-137
Rate	10 mR/hr	10.3 mR/hr	10.3 mR/hr	R-137
Rate	4 mR/hr	3.9 mR/hr	3.9 mR/hr	R-137
Rate	1 mR/hr	1.05 mR/hr	1.05 mR/hr	R-137
Integrate	200 mR	200 mR	200 mR	R-028/R-122
Integrate	50 mR	51 mR	51 mR	R-028/R-122

Readings with * indicate ranges where As-Found readings are >20% of Target value.

Readings with ** indicate As-left readings are >10% of Target value.

Readings with # indicate ranges were calibrated using a pulser

Physical Checks
Outer Physical Check: Pass
Internal Physical Check: Pass
Tap Test: Pass
Desiccant Check: Pass (Changed 1/30/25)

Comments: Dried / Replaced Desiccant.

Customer Equipment ID: 00494

Calibrated By Alana Sullivan

QC Review By Matthew Hunter

Date: Jan 30, 2025

Expires: Jan 30, 2026

QC Date: Jan 31, 2025

Atmospheric Conditions - Temperature: 70.8 °F, Humidity: 22.8 %, Barometric Pressure: 29.90 in/hg.

In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated.

Unless otherwise stated, this calibration was performed by RSCS at 93 Ledge Rd, Seabrook, NH utilizing the following NIST Traceable sources: R-028/R-122 - J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7% (Cal Due: Jan 03, 2026)

R-137 - Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0% (Cal Due: Jan 03, 2026)

Worked on calibration: Alana Sullivan

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.

Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.

Calibration Laboratory is operated in accordance with ANSI/NCCL Z540-1-1994.

RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2017 while demonstrating technical competence in the field of



Calibration Certificate
ID Number: 3096132196-0

Customer:
Upstate Medical Physics
 1290 Blossom Drive
 Victor, NY 14564

Instrument
 Fluke Biomedical Model 451P

Serial Number
 3096

Precision Check				
Test 1	Test 2	Test 3	Mean	Results
10.10 mR/hr	10.30 mR/hr	9.80 mR/hr	10.07 mR/hr	Satisfactory

Accuracy Check				
Range	Target Value	As Found	As Left	
Rate	4 R/hr	3.4 R/hr	4.1 R/hr	
Rate	1 R/hr	0.84 R/hr	0.99 R/hr	
Rate	400 mR/hr	340 mR/hr	400 mR/hr	
Rate	100 mR/hr	88 mR/hr	104 mR/hr	
Rate	40 mR/hr	37 mR/hr	39 mR/hr	
Rate	10 mR/hr	9.2 mR/hr	10.1 mR/hr	
Rate	4 mR/hr	3.2 mR/hr	3.6 mR/hr	
Rate	1 mR/hr	0.9 mR/hr	0.95 mR/hr	
Rate	400 µR/hr	370 µR/hr	370 µR/hr	
Rate	100 µR/hr	103 µR/hr	103 µR/hr	
Integrate	80 mR	69 mR	81 mR	
Integrate	20 mR	16.6 mR	19 mR	

Readings with * indicate ranges where As-Found readings are >20% of Target value. Readings with ** indicate As-left readings are >10.00% of Target value

Outer Physical Check: <i>Pass</i> Internal Check: <i>Pass</i> Tap Test: <i>Pass</i>
--

Comments: All readings 10 mR/hr or higher and the integrate readings were obtained using the J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N: 9141). All other readings were obtained using the Tech Ops Model 773 Cs-137 Beam Calibrator (S/N: S-1110).
 As-Found/As-Left Calibration Factors: F1 = 107/107, F2 = 115/120, F3 = 110/115, F4 = 100/117, F5 = 92/105, INT = 100/100

Chris Pirie
 Calibration
 Technician

QA
 Review:

Calibration Date: 04/10/2019
 Expires: 04/10/2020

Atmospheric Conditions - Temperature: 71°F Humidity: 33% Barometric Pressure: 29.64"hg

This calibration was performed by RSCS using one or more of the following NIST Traceable radiation sources:
 Tech Ops Model 773 Cs-137 Beam Calibrator (S/N S-1110), characterized using Exradin Model A6 (S/N 185) and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-008, with estimated uncertainty of 6.0%.
 J.L. Shepherd and Associates Model 89 Cs-137 Box Calibrator (S/N 9141), characterized using Exradin, Model A6 (S/N 185), A3 (S/N 197), A12 (S/N XA091124), and Keithley Electrometer Model 617 (S/N 0547677) in accordance with methods specified in RSCS TSD 11-001, with estimated uncertainty of 2.7%.
 RSCS Neutron Calibrator, AmBe Source Model NUMEC-AM-31 (S/N Am-478), characterized using Far West Technologies Model FWAD-1 "HAWK" TEPC (S/N 021) in accordance with the methods specified in RSCS TSD 13-002, with estimated uncertainty of 9.4%.
 The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%.
 Unless otherwise stated, calibrations performed in conformance to the following documents: ANSI N323AB-2013; RSCS New Hampshire Radioactive Material License Number 381R. RSCS calibration services are performed in accordance with the RSCS Radiation Protection Program Manual and Standard Operating Procedures.
 Calibration Laboratory is operated in accordance with ANSI/NCSL Z540-1-1994
 RSCS, Inc. has been assessed by ANAB and meets the requirements of ISO/IEC 17025:2005 while demonstrating technical competence in the field of calibration. Refer to the Scope of Accreditation AC-2079 for information on the types of calibrations to which this accreditation applies.
 This calibration certificate shall not be reproduced except in full without the express written consent of RSCS, Inc.





Designer and Manufacturer
of
Scientific and Industrial
Instruments

www.ludlums.com

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.

501 Oak Street
325-235-5494
Sweetwater, TX 79556, U.S.A.



CERT # 4084.01

Customer LANDAUER MEDICAL PHYSICS ORDER NO. 20440721/536615
Mfg. Fluke Model 451P-RYR Serial No. 7206
Mfg. _____ Model _____ Serial No. _____
Cal. Date 18-Feb-23 Cal Due Date 18-Feb-24 Cal. Interval 1 Year Meterface Digital

Check mark applies to applicable instr. and/or detector IAW mfg. spec. T. 76 °F RH 15 % Alt 706.6 mm Hg
 New Instrument Instrument Received Within Toler. +-10% 10-20% Out of Tol. Requiring Repair Other-See comments
 Mechanical ck. Meter Zeroed Background Subtract Input Sens. Linearity
 F/S Resp. ck Reset ck. Window Operation Geotropism
 Audio ck. Alarm Setting ck. Batt. ck.
 Calibrated in accordance with LMI SOP 14.8 Calibrated in accordance with LMI SOP 14.9

Instrument Volt Set _____ V Input Sens. _____ mV Det. Oper. _____ V at _____ mV Threshold Dial Ratio _____ = _____ mV
 HV Readout (2 points) Ref./Inst. _____ / _____ V Ref./Inst. _____ / _____ V

COMMENTS:

Instrument received with chamber depressurized.

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

Multimeter uncertainty within 1.3% of reading, Gamma uncertainty within 5.0% of reading, Neutron uncertainty within 7.0% of reading, Count rate uncertainty within 5.4% of reading

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING
Digital			

Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING	Log Scale	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING
Digital Readout						
2 R/hr	N/A	2.0 R/hr				
200 mR/hr		200 mR/hr				
20 mR/hr		20				
2 mR/hr		2.0				
200 µR/hr		200 µR/hr				

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. All pass/fail determinations are based on the manufacturer's specifications without considering uncertainty factors. Measurement results represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k=2. ISO/IEC 17025:2017[E] State of Texas Calibration License No. LO-1963 The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323AB-2013

Reference Instruments and/or Sources: Cs-137 S/N: 059 2171CP 2261CP 720 734 781 1131 1616 1696 1909 1916CP 2324/2521
 5717CO 5719CO 60646 70897 73410 E552 G112 2168CP S-394 S-1054 T10081 T10082 Neutron Am-241 Be T-304 Ra-226 Y982
 E551 5105 CSV280

Alpha S/N _____ Beta S/N _____ Other _____
 m 500 S/N _____ Oscilloscope S/N _____ Multimeter S/N _____

Calibrator Donnie Miekos / Donnie Miekos Title Calibrator Date 18 Feb 23
QC'd By Rud. L. Title Final QC Date 20 Feb 23

AC Inst. Passed Dielectric (Hi-Pot) and Continuity Test
Only Failed: _____