


Certificate number: o70059897i279575-B-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
LONG RD 334  
13159 TULLY  
US

Laboratory  
Unfors RaySafe AB  
Björklundabacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 R/F  
Serial number 279575  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2026-02-18  
As left 2026-02-28  
Adjustment done Yes  
Tested by Erik Strömberg  
Petter Olausson

Approved by

Marie Panula  
Service Technician

Certificate date

2026-03-02

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
LONG RD 334  
13159 TULLY  
US

Laboratory  
Unfors RaySafe AB  
Björklundabacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 R/F  
Serial number 279575  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2026-02-18  
As left 2026-02-28  
Adjustment done Yes  
Tested by Erik Strömberg  
Petter Olausson

Approved by

Marie Panula  
Service Technician

Certificate date

2026-03-02

## CALIBRATION RESULT

- AS FOUND:** One or more measured and tested values of this certificate are within the specifications by a margin less than the expanded measurement uncertainty. Therefore, it is not possible to state conformance with specifications for those measurements.
- AS LEFT:** All measured and tested values of this certificate were found to be in conformance with the specification.


Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %  
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o70060408i279216-B-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
LONG RD 334  
13159 TULLY  
US

Laboratory  
Unfors RaySafe AB  
Björklundabacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 279216  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2026-02-22  
As left 2026-03-02  
Adjustment done Yes  
Tested by Kristoffer Petric  
Nils Liljegren  
Pontus Ericstam

Approved by

Marie Panula  
Service Technician


Certificate date

2026-03-02

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o70060408i279216-A-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
LONG RD 334  
13159 TULLY  
US

Laboratory  
Unfors RaySafe AB  
Björklundabacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 MAM  
Serial number 279216  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2026-02-22  
As left 2026-03-02  
Adjustment done Yes  
Tested by Kristoffer Petric  
Nils Liljegen  
Pontus Ericstam

Approved by

Marie Panula  
Service Technician

Certificate date

2026-03-02

## CALIBRATION RESULT

**AS FOUND:** One or more measured and tested values of this certificate are within the specifications by a margin less than the expanded measurement uncertainty. Therefore, it is not possible to state conformance with specifications for those measurements.

**AS LEFT:** All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air Kerma: 5 %

Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
LONG RD 334  
13159 TULLY  
US

Laboratory  
Unfors RaySafe AB  
Björklundabacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com


## CUSTOMER INSTRUMENT

Product X2 CT  
Serial number 293251  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2026-02-19  
As left 2026-02-27  
Adjustment done Yes  
Tested by Erik Strömberg  
Petter Olausson

Approved by



Marie Panula  
Service Technician

Certificate date

2026-03-02

## CALIBRATION RESULT

**AS FOUND:** All measured and tested values of this certificate were found to be in conformance with the specification.

**AS LEFT:** All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
LONG RD 334  
13159 TULLY  
US

Laboratory  
Unfors RaySafe AB  
Björklundbacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 Light  
Serial number 272670  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2026-02-12  
As left 2026-02-26  
Adjustment done Yes  
Tested by Emmy Pantzar  
Approved by   
Marie Panula  
Service Technician  
Certificate date 2026-03-02

## CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

### Relevant instrument specifications

Illuminance: 5 %  
Luminance: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Certificate number: o70059388:263036-1



# Calibration Certificate



Customer  
UPSTATE MEDICAL PHYSICS  
LONG RD 334  
13159 TULLY  
US

Laboratory  
Unfors RaySafe AB  
Björklundabacken 10  
SE-436 57 Hovås  
Sweden  
+46 31 719 97 10 (phone)  
+46 31 910 950 (fax)  
customerservice.se@raysafe.com

## CUSTOMER INSTRUMENT

Product X2 Survey  
Serial number 263036  
Manufacturer RaySafe

## CALIBRATION INFORMATION

As found 2026-02-12  
As left 2026-02-26  
Adjustment done Yes  
Tested by Hussein Mahamod  
Ujjaini Joseph

Approved by

Marie Panula  
Service Technician  
2026-03-02

Certificate date

## CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

